PRESSALIT®

# United Kingdom Conformity Assessment

**United Kingdom RP:**
Pressalit Ltd.
100, Longwater Avenue
Green Park
Reading RG2 6GP
United Kingdom
T: 01180923 2696
uk@pressalit.com

This product is produced in compliance with all relevant requirements of:
- Regulation (EU) 2017/745 on medical devices (Risk class I, rule 1)
- Council Directive 2006/42/EC on machinery
- Council Directive 2011/65/EC on RoHS

For the UK:
- Medical Device Regulation (MDR) 2002 on medical devices
- Supply of Machinery (Safety) Regulations 2008
- Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

| | |
|---|---|
| Medical devices - Application of risk management and quality management in design and production | ISO 14971:2019<br>ISO 13485:2016 |

| | |
|---|---|
| The products are manufactured in accordance with the following standards: | IEC 60601-1:2005<br>ISO 17966:2016 |

Ry, 27. maj 2026

**Mette Elstrøm Kristoffersen, QA/RA Specialist, Pressalit A/S**
Responsible

**Manufacturer:** Pressalit A/S - Pressalitvej 1 DK-8580 Ry - T: +45 8788 8788 - pressalit@pressalit.com

SCT 2000

REF **R8581218000D**

| | |
|---|---|
| **EAN (GTIN):** | 5708590400303 |
| **Nomenclature (GMDN):** | 43600 |
| **Nomenclature (EMDN):** | Y093312 |

**Name**

EN SCT 2000 shower table, outer measurements 900 x 1800 mm, lying surface in mesh cover, 550 x 1650 mm, white. With safety rail and water collection tray, foldable, electrically height adjustable with wired remote control, safety stop, back-up battery, wall-mounted

**Intended purpose**

EN This product is intended for showering or changing a person who is unable to stand or sit. The person is lying horizontally on the table.
Expected service life is 10 years for institutional use, provided that maintenance is carried out as described in the user manual.