

## PRESSALIT.

| Product<br>Item number | VALUE IV back rest, fixed height<br>R1661000112   |
|------------------------|---|
| Colour                 | VVS:777525090  NRF:6311837  EAN:5708590355115  HMI: 124224  DB: 2084079  Anthracite grey  |
| Product description    | VALUE IV folding shower seat and backrest create a simple solution for users who need to sit down while showering.  VALUE IV back rest. Fixed height. Materials: Diecast aluminium and vacuum-moulded polyethylene (PE). Incl. fixing bolts for brick wall. Plugs depend on wall type and must be ordered separately. |
| Adjustability          | Fixed mounting  |
| Materials              | The wall profile is made of diecast aluminium, and the backrest is made of vacuum-moulded polyethylene (PE)   |
| Warranty               | Pressalit makes a commitment to remedy material and manufacturing defects for a period of five years from the invoicing date.   |



## PRESSALIT®

| Product                   | VALUE IV back rest, fixed height  |
|---------------------------|---|
| Item number               | R1661000112   |
| Cleaning                  | VALUE backrest The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use corrosive or abrasive cleaners as it may result in the surface being damaged. For surface disinfection, chlorine-based disinfectants may be used according to the product specification and safety precautions for these products.  The following cleaning agents have been tested by Pressalit and may be used, as long as they are diluted according to the instructions on the container.  DISINFECTANTS: Clinimax Difficil-S Erisan OXY+ Sprit (husholdningssprit) Suma Bac D10 Virkon Rodalon Desinfektionsspray Lysoformin Spezial CaviCide Sufanios Premium Tvättsprit 70% WipeClean UniversalCleaning Chrolorine Disinfection Wipes Klorilli (Kiilto Clean)  NON-DISINFECTANTS: Suma Star free D1 Taski 100 Taski 200 Milizid Citro  The product must be wiped down with a soft rag wrung out in clear water immediately after cleaning, as some materials may otherwise be discolored and stained, and the |
| Declaration of conformity | surface coating may deteriorate.  CE MDR This product is in compliance with all relevant requirements of: - Regulation (EU) 2017/745 on medical devices.  |



## PRESSALIT®

Product VALUE IV back rest, fixed height

Item number R1661000112

pressalit.com

Pressalit A/S
Pressalitvej 1
DK-8680 Ry
Tel.: +45 8788 8788
Fax: +45 8788 8789
E-mail: pressalit@pressalit.com