



United Kingdom Conformity Assessment

United Kingdom RP:
 Pressalit Ltd.
 100, Longwater Avenue
 Green Park
 Reading RG2 6GP
 United Kingdom
 T: 01180923 2696
 uk@pressalit.com

This product is produced in compliance with all relevant requirements of:
 - Regulation (EU) 2017/745 on medical devices (Risk class I, rule 1)
 - Council Directive 2006/42/EC on machinery
 - Council Directive 2011/65/EC on RoHS
 For the UK:
 - Medical Device Regulation (MDR) 2002 on medical devices
 - Supply of Machinery (Safety) Regulations 2008
 - Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Medical devices - Application of risk management and quality management in design and production	ISO 14971:2019 ISO 13485:2016
--	----------------------------------

The products are manufactured in accordance with the following standards:	IEC 60601-1:2005 ISO 17966:2016
---	------------------------------------

Ry, 27. maj 2026

Mette Elstrøm Kristoffersen, QA/RA Specialist, Pressalit A/S
 Responsible

Manufacturer: Pressalit A/S - Pressalitvej 1 DK-8580 Ry - T: +45 8788 8788 - pressalit@pressalit.com

SCT 2100	
	EAN (GTIN): 5708590376110 Nomenclature (GMDN): 43600 Nomenclature (EMDN): Y093312
	Name
	EN SCT 2100 shower table, outer measurements 900 x 1800 mm, lying surface in coated canvas, 550 x 1650 mm, white. With safety rail and water collection tray, foldable, electrically height adjustable with wired remote control, wall-mounted
Intended purpose	
EN This product is intended for showering or changing a person who is unable to stand or sit. The person is lying horizontally on the table. Expected service life is 10 years for institutional use, provided that maintenance is carried out as described in the user manual.	